



CLRD

CENTRE FOR LIVER RESEARCH AND DIAGNOSTICS

DECCAN COLLEGE OF MEDICAL
SCIENCES AND ALLIED HOSPITALS
Zafar Gadh, Kanchanbagh,
Hyderabad - 500 058. A.P., India.

College : 24340547
Owaisi Hosp. : 24341657
Tel. Fax : 24342954
E-mail: cmhabib@rediffmail.com

Dr. C.M. Habibullah

M.D., D.M. D.Sc., F.A.C.G., F.A.M.S., F.N.A.Sc.,
DIRECTOR

30th November, 2009

To,

Prof. P. V. Subba Rao
President – Research
Bigtec private Limited
Bigtec Labs, 2nd Floor
SID Entrepreneurship Centre
IISc Complex
Bangalore – 560012

Validation Certificate

Centre for Liver Diseases and Diagnostics (CLRD), Hyderabad conducted a Validation study entailing the Comparative evaluation of bigtec's microPCR against ABI 7500 for real-time PCR of Hepatitis B Virus (HBV) DNA.

All the samples were coded randomly in a double blinded manner. Based on the study the clinical specificity of bigtec's microPCR for HBV was found to be 100% and clinical sensitivity also to be 100%. The study also indicated complete absence of cross-reactivity.

CLRD certifies Bigtec's microPCR for efficient and reliable diagnosis for Hepatitis B Virus.

Truly,

(Dr.C.M.Habibullah)

Dr. C. M. HABIBULLAH
DIRECTOR

Centre For Liver Research & Diagnostic
Deccan College of Medical Sciences
HYDERABAD-500 038

Residence : "Baitul Habeeb", Venkateshwara Colony, Narayanaguda, HYDERABAD - 500 029. A.P., India.

© : 040- 24753232, Fax No. : 091 -040-24752242, Tel/fax - 24342954
website : www.clrdhyd.org