



# Truenat™

## Dengue

Chip-based Real Time PCR test for Dengue

### 1. INTENDED USE

**Truenat™ Dengue (REF 601050005 / 601050020)** is a chip-based Real Time Reverse Transcription Polymerase Chain Reaction (qRT-PCR) test for the quantitative detection and diagnosis of Dengue virus in human blood/serum/plasma specimen and aids in the diagnosis of infection with Dengue. **Truenat™ Dengue** runs on the **Truelab™ Uno** and **Truelab™ Uno Dx** Real Time micro PCR Analyzer.

### 2. INTRODUCTION

Dengue infection is a leading cause of morbidity and mortality in the tropical and subtropical regions of the world with a huge economic cost associated with it. As many as 100 million people are estimated to be infected with Dengue, yearly. Dengue infection is caused by any one of the four related serotypes of the Dengue virus (DEN-1, DEN-2, DEN-3, or DEN-4). Dengue, an RNA virus, is transmitted by the bite of an infected, day biting, *Aedes* mosquito (*Aedes aegypti* and *Aedes albopictus*) and is characterized by a sudden onset of high fever, headache, retro orbital pain in the back and limbs (break-bone fever), lymphadenopathy and maculopapular rash. Depending on the immune response, Dengue infection is classified as primary or secondary. Primary Dengue infection is usually self limiting and uncomplicated. Life threatening conditions such as Dengue Haemorrhagic Fever (DHF) and Dengue Shock Syndrome (DSS) are more severe forms of Dengue usually associated with secondary infection and is fatal if not treated early. Thus early diagnosis and treatment is crucial in the management of Dengue fever. Current methods of diagnosis of the disease include viral isolation, serologic tests and molecular methods. Serology based tests qualitatively detect IgM and IgG antibodies that appear much later during the convalescent stage or detect the NS1 antigen during the late acute phase and early convalescent phase. However, these methods are known to show cross reactivity with other flaviviruses and have limitations of sensitivity. Viral isolation and molecular techniques such as Polymerase Chain reaction (PCR) or Real Time PCR are much more sensitive and confirm infection with dengue, immediately upon onset of symptoms, during the acute (viremic) phase. However, viral isolation and PCR or Real Time PCR tests have so far been restricted to centralized reference laboratories as they require skilled manpower and elaborate infrastructure. Also the turnaround time for results could take a few days.

The **Truelab™** Real Time micro PCR System enables decentralization and near patient diagnosis of Dengue by making real time PCR technology rapid, simple, robust and user friendly and offering "sample to result" capability even at resource limited settings. This is achieved through a combination of lightweight, portable, mains / battery operated **Truelab™ Uno/ Uno Dx** Real Time micro PCR Analyzer and **Trueprep™ MAG / AUTO** Sample Prep Device and room temperature stable **Truenat™** chip-based Real Time PCR test and **Trueprep™ MAG / AUTO** Sample Prep kits so that even the peripheral laboratories with minimal infrastructure and minimally trained technician can easily perform these tests routinely in their facilities and report PCR results in less than an hour. Moreover, with these devices PCR testing can also be initiated in the field level, on site.

**Truenat™ Dengue** is a disposable, room temperature stable, chip-based Real Time PCR test with dried down PCR reagents for performing a Real Time PCR test for detection of Dengue and runs on the **Truelab™** Real Time micro PCR Analyzer. Only six (6) µL of purified RNA to be added to the reaction well for the analysis. The intelligent chip also carries test and batch related information including standard values for quantitation. The **Truenat™ Dengue** chip also stores information of used chip to prevent any accidental re-use of the chip.

**NOTE :** **Truelab™ / Truelab™ Uno / Truelab™ Uno Dx / Trueprep™ MAG / Trueprep™ AUTO Truepet™ / Truenat™** are all registered trademarks of **Molbio. Diagnostics (P) Limited**.

The **Truelab™** Real Time micro PCR Analyzer is protected by the following patents and patents pending: IN 2313/CHE/2007, WO 2009/047804 and corresponding claims of any foreign counterpart(s) thereof.

The **Truenat™** micro PCR chip is protected by the following patents and patents pending: IN 2312/CHE/2007, WO 2009/047805 and corresponding claims of any foreign counterpart(s) thereof.



### 3. PRINCIPLE OF THE TEST

**Truenat™ Dengue** works on the principle of Real Time Reverse Transcription Polymerase Chain Reaction (qRT-PCR). The RNA from the patient blood/serum/plasma sample is first extracted using **Trueprep™ MAG** Sample Prep Device and **Trueprep™ MAG** Blood Sample Prep kit or using **Trueprep™ AUTO** Universal Cartridge based Sample Prep Device and **Trueprep™ AUTO** Universal Cartridge based Sample Prep Kit. Six (6) µL of the purified RNA is then dispensed into a microtube containing freeze dried PCR reagents, including reverse transcriptase (RT). After allowing approximately 20 seconds for the dried PCR reagents to get hydrated with the RNA sample, the entire contents are pipetted out and dispensed into the reaction well of the **Truenat™ Dengue** chip. The **Truenat™ Dengue** chip is then inserted in the **Truelab™** Real Time micro PCR Analyzer where the RNA is first converted into complementary DNA (cDNA) by the RT enzyme and further thermal cycling takes place. A positive amplification causes the dual labeled fluorescent probe in the **Truenat™ Dengue** chip to release the fluorophore in an exponential manner which is then captured by the built-in opto-electronic sensor and displayed as amplification curve on the analyzer screen, on a real time basis during the test run. The Cycle threshold (Ct) is defined as the number of amplification cycles required for the fluorescent signal to cross the threshold (i.e. exceed the background signal). Ct levels are inversely proportional to the amount of target nucleic acid in the sample (i.e. the lower the Ct level the greater is the amount of target nucleic acid in the sample). Ct value is linearly correlated with amount of target RNA copies present in the sample, enabling quantitative estimation of the analyte. Standard values for every batch are present in the **Truenat™ Dengue** chip and the analyzer automatically compares these with the Ct value of the test sample to provide a quantitative result. In the case of negative samples, amplification does not occur and a horizontal amplification curve is displayed on the screen during the test run. At the end of the test run, Dengue "DETECTED" or "NOT DETECTED" result is displayed and in positive cases, Ct values and copies per milliliter (cp/ml) is displayed on the screen. Based on the Ct of the internal positive control (IPC), the validity of the test run is also displayed. The IPC is a full process control that undergoes all the processes the specimen undergoes - from extraction to amplification thereby validating the test run from sample to result. Absence of or shift of IPC Ct beyond a pre-set range in case of negative samples invalidates the test run. While IPC will co-amplify in most positive cases also, in some specimen having a high target load, the IPC may not amplify, however the test is considered valid. The results can be printed via Bluetooth using the **Truelab™** micro PCR printer or transferred to the lab computer/or any remote computer via Wifi network or 3G/GPRS network. Upto 5000 results in **Truelab™ Uno** to 20000 results in **Truelab™ Uno Dx** can be stored on the analyzer for future recall and reference.

### 4. TARGET SELECTION

The target sequence for this kit has been taken from the 3' Untranslated region (UTR) of the Dengue genome. The region selected is specific to and represents all four serotypes (DEN 1 to DEN 4) of Dengue virus.

### 5. CONTENTS OF THE Truenat™ Dengue KIT

- A. Individually sealed pouches, each containing a
  1. **Truenat™ Dengue** micro PCR chip.
  2. Microtube with freeze dried RT PCR reagents.
  3. DNase & RNase free pipette tip.
  4. Desiccant pouch.
- B. Package Insert.

REF	601050005	601050020
Σ	5 T	20 T

### 6. CONTENTS OF THE Trueprep™ AUTO Universal Sample Pre-treatment Pack (only for Trueprep™ AUTO users)

- A. Lysis buffer.
- B. Disposable transfer pipette(graduated).

REF	60205AB05	60205AB20
Σ	5T	20T

### 7. STORAGE AND STABILITY

**Truenat™ Dengue** is stable for one year from the date of manufacture if stored between 2-30°C. It is also stable for upto one (1) month at temperatures up to 40°C. Avoid exposure to light or elevated temperatures (above recommended levels).

**Trueprep™ AUTO** Universal Sample Pre-Treatment Pack is stable for 2 years from the date of manufacture if stored between 2-30°C. It is also stable for 4 weeks at temperatures upto 45°C. Do not freeze.

### 8. MATERIALS REQUIRED BUT NOT PROVIDED WITH THE KIT

**Truelab™ Real Time micro PCR Workstation** (REF 603010001/613010001/623010001) consisting of

1. **Trueprep™ MAG / AUTO** Sample Prep Device (REF 603040001/6030410001).
2. **Truelab™ Uno / Truelab™ Uno Dx** Real Time micro PCR Analyzer (REF 603020001/ 603021001).
3. **Truelab™** micro PCR Printer (REF 603050001).
4. **Truepet™** Precision Micropipettes

Also required additionally are: **Trueprep™ MAG** Blood Sample Prep Kit (REF 602010050)/**Trueprep™ AUTO** Universal Cartridge Based Sample Prep kit (REF 60203AR25), **Truenat™** Universal Control Kit (REF 601100008), DNase and RNase-free pipette tips with filter barrier, which may also be procured from **Molbio**, Powder free disposable gloves, waste disposal container with lid.

### 9. SPECIMEN PREPARATION FOR EXTRACTION WITH TRUEPREP™ MAG

**Truenat™ Dengue** requires purified nucleic acids from whole blood, serum or plasma specimen that are extracted using the **Trueprep™ MAG** Sample Prep Device and **Trueprep™ MAG** Blood Sample Prep Kit. (Refer to the User Manual of **Trueprep™ MAG** Sample Prep Device and the package insert of **Trueprep™ MAG** Blood Sample Prep Kit for details).

### 10. SPECIMEN PREPARATION FOR EXTRACTION WITH TRUEPREP™ AUTO

**Truenat™ Dengue** requires purified nucleic acids from whole blood/ plasma/serum specimen that are extracted using the **Trueprep™ AUTO** Universal Cartridge based Sample Prep Device and **Trueprep™ AUTO** Universal Cartridge based Sample Prep kit. Sample must be pre-treated using **Trueprep™ AUTO** Universal Sample Pre-treatment pack. Transfer 250µl of whole blood or 500µl of plasma/serum specimen using the transfer pipette provided into the Lysis buffer tube provided and mix well.

#### Sample Storage and Transportation:

Sample Pre-treatment decontaminates the specimen and makes it ready for storage/ transportation/ extraction. The specimen in this form is stable for up to 3 days at 40°C and 1 week at 30°C.

**Nucleic acid extraction:** Use entire content from the Lysis Buffer tube containing specimen for further procedure with the **Trueprep™ AUTO** Universal Cartridge Based Sample Prep Device and **Trueprep™ AUTO** Universal Cartridge Based Sample Prep Kit. (Refer to the User Manual of **Trueprep™ AUTO** Universal Cartridge Based Sample Prep Device and the package insert of **Trueprep™ AUTO** Universal Cartridge Based Sample Prep kit for details).

### 11. SAFETY PRECAUTIONS

1. For *in vitro* diagnostic use only.
2. Bring all reagents and specimen to room temperature (20 - 30 °C) before use.
3. Do not use kit beyond expiry date.
4. Carefully read the User Manuals and package inserts of all the components of the **Truelab™ Real Time micro PCR System** before use.
5. All materials of human origin should be handled as though potentially infectious.
6. Do not pipette any material by mouth.
7. Do not eat, drink, smoke, apply cosmetics or handle contact lenses in the area where testing is done.
8. Use protective clothing and wear disposable gloves when handling samples and while performing sample extraction.

### 12. PROCEDURAL PRECAUTIONS

1. Check all packages before using the kit. Damage to the packaging does not prevent the contents of the kit from being used. However, if the outer packaging is damaged the user must confirm that individual components of the kit are intact before using them.
2. Do not perform the test in the presence of reactive vapours (e.g., from sodium hypochlorite, acids, alkalis or aldehydes) or dust.
3. While retrieving the **Truenat™ Dengue** chip and the DNase & RNase free pipette tip from the pouch, ensure that neither bare hands nor gloves that have been used for previous tests run are used.

### 13. PROCEDURAL LIMITATIONS

1. Optimal performance of this test requires appropriate specimen collection, handling, storage and transport to the test site.
2. Though very rare, mutations within the highly conserved regions of the target genome where the **Truenat™** assay primers and/or probe bind may result in the under-quantification or a failure to detect the presence of the concerned pathogen.
3. The instruments and assay procedures are designed to minimize the risk of contamination by PCR amplification products. However, it is essential to follow good laboratory practices and ensure careful adherence to the procedures specified in this package insert for avoiding nucleic acid contamination from previous amplifications, positive controls, or specimens.
4. A specimen for which the **Truenat™** assay reports "Not Detected" cannot be concluded to be negative for the concerned pathogen. As with any diagnostic test, results from the **Truenat™** assay should be interpreted in the context of other clinical and laboratory findings.

### 14. CLEANING AND DECONTAMINATION

1. Spills of potentially infectious material should be cleaned up immediately with absorbent paper tissue and the contaminated area should be decontaminated with disinfectants such as 0.5% freshly

prepared sodium hypochlorite (10 times dilution of 5% sodium hypochlorite (household bleach) before continuing work.

- Sodium hypochlorite should not be used on an acid-containing spill unless the spill-area is wiped dry first. Materials used to clean spills, including gloves, should be disposed off as potentially bio-hazardous waste e.g. in a biohazard waste container.

## 15. TEST PROCEDURE

(Please also refer the **Truelab™ Uno/ Truelab™ Uno Dx** Real Time micro PCR Analyzer user manual)

- Switch on the **Truelab™** Analyzer.
- If using the **Truelab™ Uno** device, also switch on the touch screen. If using the **Truelab™ Uno Dx** proceed to step 3.
- Select user and enter password.
- Select the test profile for "DENGUE" on the Analyzer screen.
- Enter the patient details as prompted in the **Truelab™** Analyzer screen.
- Press Start Reaction.
- Press the eject button to open the chip tray.
- Open a pouch of **Truenat™ Dengue** and retrieve the chip-based Real Time PCR test and the microtube.
- Label the chip and the tube with the patient ID using a marker pen at the space provided on the back side of the chip and the space on the microtube label.
- Place the **Truenat™ Dengue** chip on the chip tray without touching the white reaction well. The reaction well should be facing up and away from the Analyzer. Gently press the chip to ensure that it has seated in the chip tray properly.
- Using the filter barrier tip provided in the pouch, pipette six (6) µL of the purified RNA from the Elute Collection Tube into the microtube containing the freeze dried RT-PCR reagents. Allow twenty (20) seconds for the elute to hydrate the contents. With the same tip mix the contents twice gently by pipetting in and out. Pipette out all the six (6) µL of the mixture and dispense into the centre of the white reaction well of the **Truenat™ Dengue** chip-based Real Time PCR test. Take care not to scratch the internal well surface and not to spill elute on the outside of the well.
- Slide the chip tray containing the **Truenat™ Dengue** chip-based Real Time PCR test loaded with the sample, into the **Truelab™** Analyzer.
- Press the power button on the Analyzer to turn it on. The green LED should glow.
- Press Done on the "Please Load Sample" Alert message.
- Read the result from the screen.
- Take out the **Truenat™ Dengue** micro PCR chip at the end of the test and dispose it off as per the section on "Disposal and Destruction" (Section 18).
- Turn on **Truelab™** micro PCR printer and select print on the screen for printing out hard copy of the results. Test results are automatically stored and can be retrieved any time later. (Refer to **Truelab™** Analyzer manual).
- Switch off the **Truelab™** Analyzer.

## 16. RESULTS & INTERPRETATIONS

Two amplification curves are displayed on the **Truelab™** Real Time micro PCR Analyzer screen to indicate the progress of the test. Both the target and the internal positive control (IPC)\* curves will take a steep, exponential path when the fluorescence crosses the threshold value in case of positive samples. The Ct will depend on the number of target RNA in the sample. The target curve will remain horizontal throughout the test duration and the IPC curve will take an exponential path in case of negative samples. In case the IPC curve remains horizontal in a negative sample, the test is considered as Invalid. At the end of the test run, the results screen will display "DETECTED" for Positive result or "NOT DETECTED" for Negative result. The result screen would also display the Ct value and the copies per ml (cp/ml) for positive specimen. The result screen also displays the validity of the test run as "VALID" or "INVALID". Invalid samples have to be repeated with fresh specimen from the sample preparation stage. \*While IPC will co-amplify in most positive cases also, in some specimen having a high target load, the IPC may not amplify, however the test run is still considered valid.

## 17. QUALITY CONTROL PROCEDURES

To ensure that the **Truelab™** Real Time micro PCR Analyzer is working accurately, run positive and negative controls from time to time. The **Truenat™** Universal Control Kit containing Positive Control and Negative Control must be ordered separately. It is advisable to run controls under the following circumstances: • Whenever a new shipment of test kits is received. • When opening a new test kit lot. • If the temperature of the storage area falls outside of 2-30° C. • By each new user prior to performing testing on clinical specimen.


## 18. DISPOSAL AND DESTRUCTION

- Submerge the used **Truenat™ Dengue** chip in freshly prepared 0.5% sodium hypochlorite solution for 30 minutes before disposal as per the standard medical waste disposal guidelines.
- Disinfect the solutions and/or solid waste containing biological samples before discarding them according to local regulations.
- Samples and reagents of human and animal origin, as well as contaminated materials, disposables, neutralized acids and other waste materials must be discarded according to local regulations after decontamination by immersion in a freshly prepared 0.5% of sodium hypochlorite for 30 minutes (1 volume of 5% sodium hypochlorite for 10 volumes of contaminated fluid or water).
- Do not autoclave materials or solutions containing sodium hypochlorite.
- Chemicals should be handled in accordance with Good Laboratory Practice and disposed off according to the local regulations.

## 19. SPECIFIC PERFORMANCE CHARACTERISTICS

### ASSAY RANGE AND LIMIT OF DETECTION

To determine the range of the assay, isolated RNA from standard cultures was used as source RNA. Standard curves were developed using serial dilutions of the source RNA in sterile TE buffer. The various dilutions were subjected to reverse transcription and the resulting cDNAs were tested using **Truenat™ Dengue** chip-based Real Time PCR test. The Ct values obtained were plotted against the RNA concentration. The assay was found to be linear over 5 orders of magnitude and could detect as low as 10 copies of RNA per reaction.



### CLINICAL SENSITIVITY / CLINICAL SPECIFICITY

In a prospective blinded study, 100 serum samples characterized by serology at a tertiary care hospital in Bangalore were independently tested on **Truenat™ Dengue**. Results obtained shows that out of the 100 samples **Truenat™ Dengue** detected 80 samples as positives and 20 samples as negatives, while the serology tests had reported only 60 samples as positives and

40 samples as negatives. The **Truenat™ Dengue** results were confirmed with another commercial real time PCR kit for Dengue virus detection, leading to a sensitivity and specificity of 100% for **Truenat™ Dengue**.

## REPRODUCIBILITY AND PRECISION

**Truenat™ Dengue** kit was tested for reproducibility and precision between users, analyzers and reagent lots. Serial log dilutions of a high titre clinical specimen were used for this purpose. The same sample panel was provided to 3 users, who ran it on 3 different **Truelab™ Real Time micro PCR Analyzers** with 3 different lots of **Truenat™ Dengue**. The tests were run in triplicates.

### Results of assay for reproducibility and precision:

High inter-user, inter-analyzer and inter-lot reproducibility was observed (Average standard deviation = 0.548Ct) and no significant difference was noted in the Ct values obtained from different users or analyzers or reagent lots.

## INTERFERENCE

Effect of interferences of elevated serum parameters such as lipid, cholesterol and triglycerides were evaluated in this study. Specimens (10 samples) showing higher level of serum parameters were spiked with known amounts of Dengue virus positive samples. These spiked specimen were tested using **Truenat™ Dengue**. The Ct values obtained were found not to be significantly impacted (Ct Average = 33.46; Standard Deviation = 0.31, Variance = 0.10; Co-efficient of variation = 1%) by the presence of elevated biochemical parameters as described above.

Effect of elevated biochemical blood parameters:

Sr. No.	Sample I.D.	Cholesterol	HDL	Triglycerides	LDL	Ct
1.	ST1	230	38	259	140	33.86
2.	ST2	276	35	177	194	33.73
3.	ST3	225	32	353	114	33.08
4.	ST4	304	35	394	186	33.53
5.	ST5	168	32	342	58	33.00
6.	ST6	269	46	258	176	33.22
7.	ST7	220	42	230	124	33.75
8.	ST8	94	27	56	52	33.72
9.	ST9	225	39	138	148	33.50
10.	ST10	219	24	196	128	33.21

### Carry Over Effect:

To determine whether the **Truenat™ Dengue** chip-based Real Time PCR test showed any signs of carryover of PCR products between runs, alternating runs of positive high titre samples and negatives samples were performed in duplicates. 5 positive samples and 5 negative samples were used for the study.

Results of carry-over evaluation

Sr. No.	Sample I.D.	Ct ST1	Ct ST2	Ct ST3	Average	STDEV
1.	Positive	22.55	22.67	22.6	22.61	0.06
2.	Negative	-	-	-	-	-
3.	Positive	22.5	23.2	22.3	22.67	0.47
4.	Negative	-	-	-	-	-
5.	Positive	22.7	23.5	22.6	22.93	0.49
6.	Negative	-	-	-	-	-
7.	Positive	22.78	22.89	22.7	22.79	0.10
8.	Negative	-	-	-	-	-
9.	Positive	23.4	23.1	22.5	23.00	0.46
10.	Negative	-	-	-	-	-

All the positive samples were accurately detected and all negative samples were undetected indicating absence of Carryover of PCR products between runs using the **Truenat™ Dengue** chip-based Real Time PCR test.

## 20. REFERENCES

- <http://www.cdc.gov/dengue/>. (2). <http://www.who.int/topics/dengue/en/>. (3). Vaughn DW, et al. Dengue viremia titer, antibody response pattern and virus serotype correlate with disease severity. J. Infect. Dis. 181: 2-9. (4). S A M Kularatne et al. (2009). Concurrent outbreaks of Chikungunya and Dengue fever in Kandy, Sri Lanka, 2006-07: A comparative analysis of clinical and laboratory features. Postgrad Med J 2009; 85:342-346. (5). T. Mirawati Sudiro et al. (1997) Rapid diagnosis of dengue viremia by Reverse-transcriptase Polymerase Chain Reaction using 3'-Noncoding region universal primers. Am. J. Trop. Med. Hyg., 56(4), 1997, pp. 424 - 429.

## SYMBOL KEYS

Consult instructions for use	In vitro Diagnostic Test. Not for medicinal use.	Temperature Limitation	Catalogue Number	For single use only
Manufacturer	Date of Manufacture	Date of Expiry	Batch Number / Lot Number	Contains sufficient for <n> tests
EC REP		Authorised Representative in the European Community		

**Molbio Diagnostics Pvt. Ltd.**

**Registered Office:**  
 H.No. 13, Sagar Society, Dona Paula  
 Panaji, North Goa, Goa - 403 722, India.  
[www.molbiodiagnostics.com](http://www.molbiodiagnostics.com)

**Manufacturing Unit:**  
 Plot Nos. M 46-47, Phase III B,  
 Verna Industrial Estate, Verna,  
 Goa - 403 722, India.  
 Email: [sales@molbiodiagnostics.com](mailto:sales@molbiodiagnostics.com)

EC REP

Qarad b.v.b.a. Ciplastraat 3, B-2440 Geel, Belgium