

MOLBIO DIAGNOSTICS PRIVATE LIMITED

EN ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

License No.: MFG/IVD/2021/000064

Date: 07/10/2023

Doc ID:COA/TP/014MSPT

Product Name	Trueprep® AUTO MTB Sample Pre-treatment Pack		
Generic Name	Sample pre-treament pack for MTB samples		
Lot No.	014MSPT		
Mfg. Date	2023-10		
Exp. Date	2025-09		
Pack Size	50T		
Storage condition	2°C to 40°C		

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation
Α	Contents in the carton (Trueprep® AUTO MTB Sample Pre-treatment Pack)	(Present (P)/Not Present (NP))	
1	Lysis Buffer	Present	Present
2	Liquefaction Buffer	Present	Present
3	Disposable Transfer Pipette (1 ml Graduated)	Present	Present
4	Package Insert	Present	Present

FM/QT/224-01

Page 1 of 2



	Performan	ce on Truela	b [®] Real Time micro F	PCR Analyzer	
Sr. No.	Control		Specification (Ct Values)		Test Lot (Ct Value)
	Specificity Tests	Test	IC	Test (Average)	IC (Average)
Negative (Control		2		
1	Negative MTB	ND	Control should be detected	ND	27.42
Positive C	ontrol		· ·		
1	Positive MTB	Test and Control should be detected		24.35	26.99
2	Positive MTB	Test and Control should be detected		31.57	27.58
Non Temp	olate Control			7	
1	NTC	1	ontrol should not be detected	ND	ND

*Ct- Cycle Threshold

*IC- Internal Control

*ND- Not Detected

*NTC- Non Template Control

Manufacturing records are verified, checked and product is manufactured as per the GMP, samples tested and lot is QC approved for sale purpose.

Name: Dr.Praveen Kumar MK
Department: Quality Control
Designation: QC Head

FM/QT/224-01

Page 2 of 2