

## MOLBIO DIAGNOSTICS PRIVATE LIMITED

## EN ISO 13485:2016 certified company

## **CERTIFICATE OF ANALYSIS**

License No.: MFG/IVD/2021/000064

Date: 10/10/2023

Doc ID:COA/TP/015MSPT

Product Name	Trueprep® AUTO MTB Sample Pre-treatment Pack		
Generic Name	Sample pre-treament pack for MTB samples		
Lot No.	015MSPT		
Mfg. Date	2023-10		
Exp. Date	2025-09		
Pack Size	50T, 05T		
Storage condition	2°C to 40°C		

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation
Α	Contents in the carton (Trueprep® AUTO MTB Sample Pre-treatment Pack)	(Present (P)/Not Present (NP))	
1	Lysis Buffer	Present	Present
2	Liquefaction Buffer	Present	Present
3	Disposable Transfer Pipette (1 ml Graduated)	Present	Present
4	Package Insert	Present	Present

FM/QT/224-01

Page 1 of 2



	Performan	ce on Truela	b® Real Time micro P	CR Analyzer	
Sr. No.	Control		Specification (Ct Values)		Test Lot (Ct Value)
	Specificity Tests	Test	IC	Test (Average)	IC (Average)
Negative (	Control				
1	Negative MTB	ND	Control should be detected	ND	28.67
Positive C	ontrol				
1	Positive MTB	Test and Control should be detected		27.87	28.02
2	Positive MTB	Test and Control should be detected		29.68	27.64
Non Temp	olate Control				
1	NTC	I COMMITTED TO SERVICE STATES	ontrol should not be detected	ND	ND

\*Ct- Cycle Threshold

\*IC- Internal Control

\*ND- Not Detected

\*NTC- Non Template Control

Manufacturing records are verified, checked and product is manufactured as per the GMP, samples tested and lot is QC approved for sale purpose.

Name: Dr.Praveen Kumar MK Department: Quality Control Designation: QC Head