

MOLBIO DIAGNOSTICS PRIVATE LIMITED

EN ISO 13485:2016 certified company

CERTIFICATE OF ANALYSIS

License No.: MFG/IVD/2019/000003

Date: 11/06/2024

Doc ID:COA/TP/MSPT219/50T

Product Name	Trueprep® AUTO MTB Sample Pre-treatment Pack	
Generic Name	Sample pre-treament pack for MTB samples	
Lot No.	MSPT219	
Mfg. Date	2024-06	
Exp. Date	2026-05	
Pack Size	50T	
Storage condition	2°C to 40°C	

PRECAUTION: Refer Safety Data Sheet and follow instructions for use.

Sr. No.	Components in each kit	Specification	Observation
Α .	Contents in the carton (Trueprep® AUTO MTB Sample Pre-treatment Pack)	(Present (P)/Not Present (NP))	
1	Lysis Buffer	Present	Present
2	Liquefaction Buffer	Present	Present
3	Disposable Transfer Pipette (1 ml Graduated)	Present	Present
4	Package Insert	Present	Present



	Performan	ce on Truelat	o® Real Time micro	PCR Analyzer	
Sr. No.	Control		Specification (Ct Values)		Test Lot (Ct Value)
	Specificity Tests	Test	IC	Test (Average)	IC (Average)
Negative	Control				
1	Negative MTB	ND	Control should be detected	ND	27.58
Positive (Control				.
1	Positive MTB	Test and Control should be detected		30.26	26.13
2	Positive MTB	Test and Control should be detected		24.54	27.58
Non Tem	plate Control				
1	NTC	Test and Control should not be detected		ND	ND

^{*}Ct- Cycle Threshold

Manufacturing records are verified, checked and product is manufactured as per the GMP, samples tested and lot is QC approved for sale purpose.

> Name: Dr.Praveen Kumar M **Department:** Quality Control Designation: QC Head

FM/QT/224-01

^{*}IC- Internal Control *ND- Not Detected

^{*}NTC- Non Template Control