

## MOLBIO DIAGNOSTICS PRIVATE LIMITED

## EN ISO 13485:2016 certified company

## **CERTIFICATE OF ANALYSIS**

License No.: MFG/IVD/2019/000003

Date: 19/06/2024

Doc ID:COA/TP/MSPT221/05T

Product Name	Trueprep® AUTO MTB Sample Pre-treatment Pack		
Generic Name	ic Name Sample pre-treament pack for MTB samples		
Lot No.	MSPT221		
Mfg. Date	2024-06		
Exp. Date	2026-05		
Pack Size	05T		
Storage condition	2°C to 40°C		

PRECAUTION: Refer Safety Data Sheet and follow instructions for use

Sr. No.	Components in each kit	Specification	Observation
Α	Contents in the carton (Trueprep® AUTO MTB Sample Pre-treatment Pack)	(Present (P)/Not Present (NP))	
1	Lysis Buffer	Present	Present
2	Liquefaction Buffer	Present	Present
3	Disposable Transfer Pipette (1 ml Graduated)	Present	Present
4	Package Insert	Present	Present



×1	Performand	e on Truelab	® Real Time micro	PCR Analyzer	
Sr. No.	Control		Specification (Ct Values)		Test Lot (Ct Value)
	Specificity Tests	Test	IC	Test (Average)	IC (Average)
Negative	Control				
1	Negative MTB	ND	Control should be detected	ND	26.72
Positive	Control				
1	Positive MTB	Test and Control should be detected		31.08	28.12
2	Positive MTB	Test and Control should be detected		24.76	26.96
Non Tem	plate Control				
1	NTC	Test and Control should not be detected		ND	ND

<sup>\*</sup>Ct- Cycle Threshold

Manufacturing records are verified, checked and product is manufactured as per the GMP, samples tested and lot is QC approved for sale purpose.

Name: Dr.Praveen Kumar MK
Department: Quality Control
Designation: QC Head

<sup>\*</sup>IC- Internal Control

<sup>\*</sup>ND- Not Detected

<sup>\*</sup>NTC- Non Template Control